

Certificate of Analysis

This lot of product meets the accepted performance criteria recommended in ISO 11140-1: 2014 as a Type 4 Multi-critical Process Variable Indicator. Product Name: Indicator for Monitoring Low Temperature Steam Formaldehyde Processes REF LOT F181025 EFVG-14 2021-10-25 (YYYY-MM-DD) M 2018-10-25 Steam Formaldehyde For use in Monitoring: Appearance: Indicator Strip-Label, 19 mm x 64 mm Performance Characteristics Characteristic Specification Result Steam Formaldehyde Violet to Green Color Transition Pass 70°C, 15 minutes, 1.0 mol/L Date of Product Performance Verification: October 2018 Storage and Shelf Life **−**+30°C Keep away from sunlight Room Temperature (15° – 30°C) Protect from heat, radioactive sources, and 20% to 80% Relative Humidity sterilizing agents Shelf Life 36 Months from the date of manufacture Disposal: Discard as general waste.

Date

True Indicating, LLC 946 Kane Street, Suite A Toledo, Ohio 43612

Quality Approval

www.trueindicating.com

info@trueindicating.com