



Technical Data Sheet

PREPARED MODIFIED GROWTH MEDIUM For Use With *Bacillus athrophaeus*

True Indicating Codes: GGM-100 and GGM-50R



Product Description

The prepared growth media consists of an exclusively formulated Trypticase Soybean Broth (TSB) modified with pH indicator. The pH indicator causes a transition in media color from Green to Yellow when *Bacillus athrophaeus* bacterial growth is present, allowing for a reduced incubation when used in conjunction with True Indicating's Biological Indicators (BIs) and Inoculated Carriers.

Physical Properties

Tube Dimensions	16.7mm x 62.8mm
Fill Volume	6.5 mL \pm 0. 5mL
Packaging	GGM-100: 100 Tubes / Pack GGM-50R: 50 Tubes + Transport Rack

Indications for Use

The GGM growth medium is designed to be utilized in conjunction with BIs containing *Bacillus athrophaeus* to monitor Ethylene Oxide (EO), Dry Heat or Vaporised Hydrogen Peroxide (VH₂O₂) processes.

Monitoring Frequency

For greatest control of sterilized goods, it is recommended that one or more BIs be included with every load. To maximize efficiency and reduce incubation time, use modified growth medium in conjunction with

Instructions for Use

Transfer: Using aseptic technique, typically in a laminar flow hood, transfer one processed (exposed) BI or inoculated carrier to each tube of growth medium.

Controls: A tube of growth medium, without an inoculated carrier, may be incubated as outlined below as a negative control. Label the tube as Negative Control.

When a positive control is needed, aseptically transfer an unprocessed (unexposed) inoculated carrier into a tube of growth medium. Label the tube as Positive Control.

Incubation: Place the tube(s) in a vertical position in an incubator set at 30°C to 40°C.

Sterilization Process	Minimum Incubation Time
EO or Dry Heat	48 Hours
VH ₂ O ₂	72 Hours



TRUE INDICATING
PRODUCTS

946 Kane St. Suite A, Toledo, OH, 43612
United States
@: info@trueindicating .com
www.trueindicating.com



Technical Data Sheet

Monitoring: Examine the tubes daily, whenever possible during incubation as outlined below and record observations.

Interpretation:

Negative Control: The Negative Control should exhibit a Green color and should remain clear with no signs of turbidity. If the control transitions to Yellow or shows signs of growth, consider the test invalid.








Positive Control: The Positive Control should exhibit a color change to Yellow and/or demonstrate turbidity. If the positive control does not show signs of growth, consider the test invalid.

Test: A passing cycle is indicated by the color of the growth medium remaining Green and the media remaining clear with no signs of turbidity. Failed cycle is indicated by a color change to Yellow and the /or presence of turbidity.

Performance Characteristics

Sterility	Pass
Growth Promotion Capabilities	Growth of <i>Bacillus atrophaeus</i> cell line 9372 within 48 hours

Storage and Shelf Life

	15°C to 30°C		Keep away from Sunlight
	20% to 80% Relative Humidity		Keep dry
	Do not freeze		Do not use after expiration date
Shelf-Life	12 Months from the date of manufacture		Protect from heat & radioactive sources

Disposal

Prior to disposal, Autoclave all positive units at 121°C for not less than 30 minutes. Recycle glass tubes whenever possible.

